

BIOCLEAN STERILE NITRILE High Grip RABS / Isolator Gloves GHG

About

BioClean validated sterile RABS and Isolator Gloves are manufactured from nitrile with incredibly low levels of particles and excellent ESD properties. Designed for precision work when increased grip is required, our GHG series of high grip gloves are fully validated for sterility with an SAL (Sterility Assurance Level) of 10^{-6} and are available in a range of port sizes.

Specifications

COMPATIBILITY:

ISO Class 4 & EU GMP Grade A

LENGTH:

840mm/33"

MATERIAL:

Nitrile

PROTEIN LEVEL:

Nitrile contains no natural latex proteins

SURFACE:

Smooth

SHAPE:

5 Finger Ambidextrous Size 9.75

COLOUR:

White

STERILIZATION:

Gamma irradiation

Features

- Tested against ASTM D6978-05 for handling chemo drugs
- Ultra-clean surface ensures product protection
- Fits 4 port sizes:
 - GHG15NIT59 - 6-8" / 152-203mm
 - GHG20NIT59 - 8-10" / 203-254mm
 - GHG33NIT59 - 10-12" / 254-305mm
 - GHG36NIT59 - 12-14" / 305-356mm

100% INSPECTED AND AIR LEAK TESTED


SUITABLE FOR AUTOCLAVING

CAN BE SANITIZED BY VHP OR IPA

HIGH GRIP



EN 421:2010 ISO 374-5:2016



Ansell

Quality Standards

- Manufactured in a facility holding ISO 9001:2015 and PPER 2016/425 Module D
- Conforms to Category 3 - Complex Design Personal Protective Equipment PPE Regulation (EU)2016/425
- Processed in an NEBB certified ISO Class 4 environment
- Complies with the requirements of Glove Standard EN420:2003+A1:2009, EN374-1:2016, EN374-5:2016 and EN421:2010
- Physical properties comply with European Medical Glove Standard EN455-2:2015
- Tested against permeation standard ASTM D6978-05 for cytotoxic drugs
- Suitable for contact with dry, alcoholic and acidic foodstuffs . Regulation (EC)1935/2004 and (EC) 2023/2006
- Sterilized in accordance with ISO 11137-1:2006
- Mechanical protection EN388:2016
- Complies with the requirements of European Glove Standard EN421 Protection against particulate radioactive contamination.

Physical Properties

CHARACTERISTICS	VALUE	TEST METHOD
Freedom From Holes	100% Tested	EN 374-2:2016
Glove Length(mm/in) (-10mm/+50mm)	840/33	EN 420:2003 + A1:2009
Single Wall Thickness (mils/mm) (± 2 mil/0.05 mm)	Cuff: 12/0.30 Palm: 20/0.50 Finger Tip: 24/0.60	EN 455-2:2015
Force At Break During Shelf Life	≥ 9N	EN 455-2:2015

PALM WIDTHS (mm) ± 5mm				
SIZE	9.75	9.75	9.75	9.75
mm	127	127	127	127
TEST METHOD	EN 420:2003 + A1:2009			

Sterilization

Method: Gamma irradiation
Minimum Dose: 25kGy
Sterility Assurance Level: 10⁻⁶

Shelf Life & Storage

Three (3) years from date of manufacture. Store in a dry, cool place (<40°C) away from direct sunlight and fluorescent light.



Protection Against Mechanical Risks, EN 388: 2016+A1:2018

Risk	Performance Level	Risk	Performance
Abrasion Resistance	Level 4	Tear resistance	Level 0
Blade cut resistance	Level 1	Puncture resistance	Level 1
TDM: Cut Resistance	X	Impact Protection	N/A

TO ORDER

Re-order code

GHG15NIT59
GHG20NIT59
GHG33NIT59
GHG36NIT59

Size

9.75
9.75
9.75
9.75

Packing

Triple bagged: One piece per sealed inner PE bag; one inner bag per sealed second inner PE bag; one second inner bag per sealed outer PE bag; 20 outer bags per lined inner white Correx polyethylene box (20 pieces).

For additional information visit us at www.bioclean.com, or contact us at;

Nitritex Ltd

Minton Enterprise Park
Oaks Drive
Newmarket, Suffolk
CB8 7YY
United Kingdom
Tel: +44 1638 663338

Ansell US

111 Wood Avenue South
Suite 210
Iselin, NJ 08830
USA
Tel: +1 800 800 0444

Nitritex (M) Sdn.Bhd

2 Jalan Jurunilai U1/20
Seksyen U1, Hicom Glenmarie Industrial Park
40150 Shah Alam, Selangor
Malaysia
Tel: +603 5569 3857/3859

GHG/1906/PDS11



Neither this document nor any other statement made herein by or on behalf of Ansell should be construed as a warranty of merchantability or that any Ansell product is fit for a particular purpose. Ansell assumes no responsibility for the suitability or adequacy of an end user's selection of gloves for a specific application.

*Warning: No glove provides complete protection against all chemicals. Users must test the gloves against the particular chemicals and environment in which they will be used.

**Please see product validation pack or contact Ansell customer service for specific data on use of gloves with cytotoxic drugs. Gloves used for protection against such drugs must be selected specifically for the type of chemicals used.

Ansell, ® and TM are trademarks owned by Ansell Limited or one of its affiliates. © 2012 – 2019 Ansell Limited. All Rights Reserved.