Riverside Business Park – Block J Boulevard International 55 1070 Brussels, Belgium T. + 32 (0)2 528 74 00 F. + 32 (0)2 528 74 01

Email: info.europe@ansell.com www.ansell.com



EU DECLARATION OF CONFORMITY

This EU Declaration of Conformity is issued under the sole responsibility of Ansell Healthcare Europe NV.

Manufacturer Name/Address: Ansell Healthcare Europe NV

Boulevard International 55

Brussels B-1070 Belgium

SRN Number: BE-MF-000000691

Risk Class: Class IIa

Intended Purpose: A sterile medical device intended as a surgical glove and

a protective barrier when worn on the hands of healthcare providers at the surgical site. It is used mainly as a two-way barrier to protect patient and staff from

microorganisms. This is a single-use device.

EMDN Code and Description: T01010102 - Non-Powdered Latex Surgical Gloves

Basic UDI DI: 5414566 GLM33004 HS

Product Name(s):

Product Name	Product Code	Size	Region(s)
Gammex® Latex Moisturizing	330049055	5.5	EMEA
Gammex® Latex Moisturizing	330049060	6	EMEA
Gammex® Latex Moisturizing	330049065	6.5	EMEA
Gammex® Latex Moisturizing	330049070	7	EMEA
Gammex® Latex Moisturizing	330049075	7.5	EMEA
Gammex® Latex Moisturizing	330049080	8	EMEA
Gammex® Latex Moisturizing	330049085	8.5	EMEA
Gammex® Latex Moisturizing	330049090	9	EMEA

Conformity Assessment Procedure: Annex IX.

CE Certificate No: MDR 763361.

Certified through the British Standards Institution, Notified Body Number 2797.

Ansell Healthcare Europe NV/SA

Riverside Business Park – Block J Boulevard International 55 1070 Brussels, Belgium T. + 32 (0)2 528 74 00 F. + 32 (0)2 528 74 01

Email: info.europe@ansell.com www.ansell.com



Ansell Healthcare Europe NV Riverside Business Park - Block J Bld Internationalelaan 55 B-1070 Brussels BELGIUM

We hereby declare that the medical device(s) specified above meet the provision of the Regulation (EU) MDR 2017/745 for medical devices.

Signed on behalf of the Manufacturer

Name: Samantha Marshall

Position: Director Regulatory Affairs Medical EMEA & APAC

Date: 02 March 2023 Place of Issue: Nuneaton, England

Revision: MED\MDR\GLMOISTUR\001