

## **EU DECLARATION OF CONFORMITY**

**Manufacturer Name/Address:** Ansell Healthcare Europe NV/SA  
Riverside Business Park,  
Block J, Boulevard International 55,  
1070 Brussels,  
Belgium

**SRN Number:** BE-MF-000000691

**Risk Class:** Class I

**Intended Purpose:** A non-sterile medical device intended as an examination/treatment glove and a protective barrier when worn on the hands of healthcare providers, during patient examination, or for other sanitary purposes. The device is used mainly as a two-way barrier to protect both the patient and wearer against contaminants. This is a single-use device.

**EMDN Code and Description:** T01020204 - Nitrile Examination/Treatment Glove

**Basic UDI-DI:** 5414566 MTN30 WG

**Product Name(s):**

Product Name	Size	Product Code	Market Regions
Micro-Touch® N30	XS	313043060	EMEA/Saudi Arabia
Micro-Touch® N30	S	313043065	EMEA/Saudi Arabia
Micro-Touch® N30	M	313043070	EMEA/Saudi Arabia
Micro-Touch® N30	L	313043075	EMEA/Saudi Arabia
Micro-Touch® N30	XL	313043080	EMEA/Saudi Arabia
Micro-Touch® N30	XS	700291	APAC/India
Micro-Touch® N30	S	700292	APAC/India
Micro-Touch® N30	M	700293	APAC/India
Micro-Touch® N30	L	700294	APAC/India
Micro-Touch® N30	XL	700295	APAC/India

**Conformity Assessment Procedure:** Annex I & Annex II + Annex III

This EU Declaration of Conformity is issued under the sole responsibility of Ansell Healthcare Europe NV/SA.

We hereby declare that the medical device(s) specified above meet the provision of the Regulation (EU) MDR 2017/745 for medical devices.

Signed on behalf of Ansell Healthcare Europe NV



Ansell Healthcare Europe NV  
Riverside Business Park - Block J  
Bld Internationalelaan 55  
B-1070 Brussels  
BELGIUM

Name: Samantha Marshall  
Position: Director Regulatory Affairs Medical EMEA / APAC  
Date of issue: 02 March 2023  
Place of issue: Nuneaton, England  
Version No: MED\MTN30\003