

# UK DECLARATION OF CONFORMITY

The Manufacturer  
**ANSELL HEALTHCARE EUROPE N.V.**  
**RIVERSIDE BUSINESS PARK, BLOCK J**  
**BOULEVARD INTERNATIONAL 55**  
**B-1070 BRUSSELS**  
**BELGIUM**

declares under his sole responsibility, that the PPE described hereafter:

## **Micro-Touch<sup>®</sup> Nitra-Free<sup>™</sup>** **PPE to be used against category III risks**



EN ISO 374-1/Type C



EN ISO 374-5



VIRUS

is in conformity with the provisions of Regulation 2016/425 on personal protective equipment, as amended to apply in GB, and with the standards EN 421:2010, EN ISO 21420:2020, EN ISO 374-1:2016, EN ISO 374-5:2016 and is identical to the PPE which is subject to the UK Type-examination (Module B, Annex V of the Regulation), under certificate number AB0321/22924-01/E00-00 , issued by the Approved Body:

**SATRA TECHNOLOGY CENTRE**  
**WYNDHAM WAY, TELFORD WAY,**  
**KETTERING, NORTHAMPTONSHIRE,**  
**NN16 8SD, UNITED KINGDOM**

and is subject to the conformity assessment procedure set out in Annex VII (Module C2) of the Regulation under the surveillance of the Approved Body:

**SATRA TECHNOLOGY CENTRE**  
**WYNDHAM WAY, TELFORD WAY,**  
**KETTERING, NORTHAMPTONSHIRE,**  
**NN16 8SD, UNITED KINGDOM**

**Guido Van Duren**  
**Director - Regulatory affairs**  
**Ansell**

**Place: Brussels**  
**Date: 2022/11/24**