

High-visibility orange nitrile exam gloves, for protection that you can see

- **High visibility:** MICROFLEX™ 93-856 are robust, high visibility orange nitrile medical exam gloves that protect hands in demanding environments — their high-visibility orange color makes them easy to see in emergency situations
- **Secure grip:** These gloves work exceptionally well in oily environments — with their textured fingers, wearers can confidently grip tools and materials in every situation
- **Added protection:** These orange nitrile gloves also feature an extended cuff for added protection of the wrist and forearm, and offer additional protection against chemotherapy drugs*



*see product packaging for details

Industries

- Automotive
- Automotive Aftermarket
- Life Sciences
- Mining
- Energy
- Healthcare

Recommended For

- General Purpose Auto Aftermarket
- Sample taking and processing
- Raw material sample collection
- Product inspection
- Vehicle maintenance
- Protection from Type I latex allergy in HCW's or patients
- Picking parts and accessories
- Handling of tools
- Inspection, selecting, checking parts
- Assembly and inspection of components
- Equipment repair and maintenance
- Administering drugs
- Inspection of primary product
- Extra protection over the wrist and arm

TECHNICAL DATA SHEET

Product Information	
Material	Nitrile
External Glove Surface	Textured Fingers
Audit Standards	ISO 13485:2003
Product Certification	Medical Device Regulation (EU) 2017/745 Class I, Personal Protective Equipment Regulation (EU) 2016/425 Category III risks, CE 0493, European Directive 89/686/EEC, EN 455 1-4, EN ISO 374-1:2016 Type B JKT, EN ISO 374-5:2016 Virus, FDA21 CFR 177-2600-US Food Contact Approved, UKCA
Packaging Overview	100 gloves per dispenser 10 dispensers per case 1000 gloves per case
Storage Instructions	Keep out of direct sunlight; store in a cool and dry place. Keep away from sources of ozone or ignition.
Country Of Origin	Malaysia
Product Segmentation	High Risk
Antistatic	Yes EN1149
Silicone Free	No
Vulcanization Chemical Accelerators	Zinc Dibutyldithiocarbamate (ZDBC) <i>Only a very small number of users may be sensitive to this ingredient(s) and hence may develop irritant and/or allergic contact reactions.</i>

PHYSICAL PROPERTIES

	Typical Values		Testing Method
Length (mm/inches)	270 / 10.6		ASTM D3767,EN 420
Freedom from Holes (Inspection level I)	1.5 AQL		ASTM D5151,EN 455-1
Palm Thickness (mm/mils)	0.13 / 5.1		ASTM D3767,EN 420
Finger Thickness (mm/mils)	0.20 / 7.9		ASTM D3767,EN 420
	BEFORE AGING	AFTER AGING	
Elongation at Break (%)	≥ 500	≥ 400	ASTM D412
Force at break (N)	≥9	≥10	EN 455-2

ORDERING INFORMATION

Size	S (6.5 - 7)	M (7.5 - 8)	L (8.5 - 9)	XL (9.5 - 10)	XXL (10.5 - 11)
Product Code	93856070	93856080	93856090	93856100	93856110

Performance Standards and Regulatory Compliance

CE0493

MD

EN ISO 374-1:2016
Type B



J K T

EN ISO 374-5



VIRUS

UK
CA

For additional information visit us at www.ansell.com, or call us at

Europe, Middle East & Africa Region
Ansell Healthcare Europe NV
T: +32 (0) 2 528 74 00
F: +32 (0) 2 528 74 01

Asia Pacific Region
MY: +603 8310 6688
apac.medical@ansell.com
CN: +86 21 38275000
infochina@ansell.com
JP: +813 5549 8151
info.medical.jp@ansell.com

North America Region
Ansell Healthcare Products LLC
T: +1 800 800 0444
F: +1 800 800 0445

Latin America & Caribbean Region
BR: +55 (11) 3356-3100
latam.medical@ansell.com
MX: +52 442 296 2050
latam.medical@ansell.com
CO: +57 1 288 3247
latam.medical@ansell.com

Australia
Ansell Limited
T: +61 1800 337 041
F: +61 1800 803 578

Russia
Ansell PYC
Ten. +7 495 258 13 16

Ansell, ® and ™ are trademarks owned by Ansell Limited or one of its affiliates. US Patented and US and non-US Patents Pending: www.ansell.com/patentmarking © 2025 Ansell Limited. All Rights Reserved.

Neither this document nor any other statement made herein by or on behalf of Ansell should be construed as a warranty of merchantability or that any Ansell product is fit for a particular purpose. Ansell assumes no responsibility for the suitability or adequacy of an end user's selection of gloves for a specific application.

