

Black nitrile exam gloves, delivering reliable extended protection with secure grip

- **Extended protection:** A longer cuff offers expanded protection of the forearm and wrist
- **Dark black color:** Dirt-masking black color makes potential hazards easier to spot
- **Elevated comfort and grip:** The soft, flexible nitrile composition provides comfort while the textured design ensures a confident grip
- **Certified performance:** These gloves meet National Fire Protection Association 1950 Emergency Medical Glove Requirements, indicating they are approved for use by first responders in the US



Industries

- Automotive
- Automotive Aftermarket
- Correctional Facilities
- Emergency Medical Services
- Fire and Rescue
- Life Sciences
- Chemical
- Manufacturing
- Recycling and Waste Disposal
- Transportation

Recommended For

- General Purpose Auto Aftermarket
- Raw material sample collection
- Overdose situations
- Laboratory analysis
- Paint shop
- Positioning patients
- Oil, fluids and filter change
- Administering drugs
- Equipment repair and maintenance
- Extra protection over the wrist and arm
- Standard patient examinations
- Emergency services
- Inspection of parts, equipment

TECHNICAL DATA SHEET

Product Information	
Material	Nitrile
External Glove Surface	Fully Textured
Audit Standards	ISO 13485, ISO 9001
Product Certification	Medical Device Regulation (EU) 2017/745 Class I, Personal Protective Equipment Regulation (EU) 2016/425 Category III risks, ASTM D6319, CE 0493, EC 1935/2004, EC 2002/72, EC 2023/2006, EN 455 1-4, EN ISO 21420:2020, EN ISO 374-1:2016 Type B JKPT, EN ISO 374-5:2016 Virus, EN 421:2010 Radioactive Contamination, FDA21 CFR 177-2600-US Food Contact Approved, NFPA 1999:2018, ISO 9001, ISO 10993-10, ISO 10993-23, ISO 13485, ISO 18889:2019, UKCA, EU Food Contact Approved
Packaging Overview	100 gloves per dispenser 10 dispensers per case 1000 gloves per case
Storage Instructions	Keep out of direct sunlight; store in a cool and dry place. Keep away from sources of ozone or ignition.
Country Of Origin	Thailand
Product Segmentation	High Risk
Antistatic	Yes EN1149
Silicone Free	No
Vulcanization Chemical Accelerators	- Zinc Diethyldithiocarbamate (ZDEC) - Zinc Dibutyldithiocarbamate (ZDBC) <i>Only a very small number of users may be sensitive to this ingredient(s) and hence may develop irritant and/or allergic contact reactions.</i>

PHYSICAL PROPERTIES

	Typical Values		Testing Method
Length (mm/inches)	300 / 12		ASTM D3767,EN 420
Freedom from Holes (Inspection level I)	0.65 AQL		ASTM D5151,EN 374-2
Palm Thickness (mm/mils)	0.15/5.9		ASTM D3767,EN 420
Finger Thickness (mm/mils)	0.17/6.7		ASTM D3767,EN 420
	BEFORE AGING	AFTER AGING	
Elongation at Break (%)	500	500	ASTM D412
Force at break (N)	6	6	EN 455-2

ORDERING INFORMATION

Size	S (6.5 - 7)	M (7.5 - 8)	L (8.5 - 9)	XL (9.5 - 10)	XXL (10.5 - 11)
Product Code	93862070	93862080	93862090	93862100	93862110

Performance Standards and Regulatory Compliance



For additional information visit us at www.ansell.com, or call us at

Europe, Middle East & Africa Region Ansell Healthcare Europe NV T: +32 (0) 2 528 74 00 F: +32 (0) 2 528 74 01 Asia Pacific Region MY: +603 8310 6688 apac.medical@ansell.com CN: +86 21 38275000 infochina@ansell.com JP: +813 5549 8151 info.medical.jp@ansell.com	North America Region Ansell Healthcare Products LLC T: +1 800 800 0444 F: +1 800 800 0445 Latin America & Caribbean Region BR: +55 (11) 3356-3100 latam.medical@ansell.com MX: +52 442 296 2050 latam.medical@ansell.com CO: +57 1 288 3247 latam.medical@ansell.com	Australia Ansell Limited T: +61 1800 337 041 F: +61 1800 803 578 Russia Ansell PYC Ten. +7 495 258 13 16
---	--	--

Ansell, ® and ™ are trademarks owned by Ansell Limited or one of its affiliates. US Patented and US and non-US Patents Pending: www.ansell.com/patentmarking © 2025 Ansell Limited. All Rights Reserved.

Neither this document nor any other statement made herein by or on behalf of Ansell should be construed as a warranty of merchantability or that any Ansell product is fit for a particular purpose. Ansell assumes no responsibility for the suitability or adequacy of an end user's selection of gloves for a specific application.

