

# UK DECLARATION OF CONFORMITY

The Manufacturer  
**ANSELL HEALTHCARE EUROPE N.V.**  
**RIVERSIDE BUSINESS PARK, BLOCK J**  
**BOULEVARD INTERNATIONAL 55**  
**B-1070 BRUSSELS**  
**BELGIUM**  
**WWW.ANSELL.COM**

UK Importer  
**PATIENT GUARD LTD**  
**LANCASTER HOUSE,**  
**AMY JOHNSON WAY,**  
**BLACKPOOL, LANCASHIRE,**  
**FY4 2RP, UNITED KINGDOM**  
**INFO@PATIENTGUARD.CO.UK**

declare under their sole responsibility, that the PPE described hereafter:

## **MICRO-TOUCH™ Nitrile Sterile XTRA**

**PPE to be used against category III risks**

EN ISO 374-1:2016  
Type B



**K P T**

EN ISO 374-5



**VIRUS**

is in conformity with the provisions of Regulation 2016/425 on personal protective equipment, as amended to apply in GB, and with the standards EN ISO 21420:2020, EN ISO 374-1:2016+A1:2018, EN ISO 374-5:2016 and is identical to the PPE which is subject to the Type-examination (Module B, Annex V of the Regulation), under certificate number 032/2025/0626, issued by the Body:

**CENTEXBEL (0493)**  
**TECHNOLOGIEPARK 70**  
**B-9052 ZWIJNAARDE**  
**BELGIUM**

and is subject to the conformity assessment procedure set in out in Annex VIII (Module D) of the Regulation under the surveillance of the Body:

**CENTEXBEL (0493)**  
**TECHNOLOGIEPARK 70**  
**B-9052 ZWIJNAARDE**  
**BELGIUM**

**Guido Van Duren**  
**Director - Regulatory affairs**  
**Ansell**

**Place: Brussels**  
**Date: 2025/10/23**