

Medical exam gloves with extended chemical splash protection, approved for handling chemotherapy drugs

- **Strong formulation:** MICROFLEX™ 93-853 exam gloves' robust nitrile formulation makes them more durable and resistant to rips, snags and tears
- **Specialized protection:** Simulating real-world overdose conditions, these gloves are specially tested against gastric acid (vomit) and fentanyl, as well as cytotoxic/chemotherapy drugs*
- **Real-world performance:** These sturdy medical exam gloves also satisfy NFPA 1999 Emergency Medical Glove Requirements, making them suitable for healthcare professionals, first responders and lab workers
- **Low AQL:** These medical exam gloves' low 0.65 AQL reinforces their barrier integrity
- **Extended protection:** Featuring an extended cuff, these gloves offer forearm and wrist protection for medical workers



*see product packaging for details

Industries

- Automotive
- Life Sciences
- Healthcare
- Machinery and Equipment

Recommended For

- General Purpose Auto Aftermarket
- Tire rotation & changes
- Sample taking and processing
- Raw material sample collection
- Maintaining lab furniture and equipment
- Inspection, selecting, checking parts
- Assembly and inspection of components
- Picking and fastening components
- Equipment repair and maintenance
- Animal care feeding and cleanup
- Administering drugs

TECHNICAL DATA SHEET

Product Information	
Material	Nitrile
External Glove Surface	Textured Fingers
Audit Standards	ISO 13485, ISO 9001
Product Certification	Medical Device Regulation (EU) 2017/745 Class I, Personal Protective Equipment Regulation (EU) 2016/425 Category III risks, CE 0493, EN 1149-1/2/3, EN 16523-1, EN ISO 374-1:2016 Type B KPT, EN ISO 374-5:2016 Virus, EN 455 1-4, EN ISO 21420:2020, ISO 11193, ISO 18889:2019, PMDA Notification Number: 13B3X10307AH0001, Trade Name: Ansell Non Latex Medical Examination Glove, UKCA,EU Food Contact Approved
Packaging Overview	50 gloves per dispenser 10 dispensers per case 500 gloves per case 40 gloves per dispenser (Size XXXL) 400 gloves per case (Size XXXL)
Storage Instructions	Keep out of direct sunlight; store in a cool and dry place. Keep away from sources of ozone or ignition.
Country Of Origin	Malaysia
Product Segmentation	High Risk
Antistatic	Yes EN1149
Silicone Free	No

PHYSICAL PROPERTIES

	Typical Values		Testing Method
Length (mm/inches)	295 / 11.6		ASTM D3767,EN 420
Freedom from Holes (Inspection level I)	0.65 AQL		ASTM D5151,EN 455-1
Palm Thickness (mm/mils)	0.14/5.5		ASTM D3767,EN 420
Finger Thickness (mm/mils)	0.22/8.7		ASTM D3767,EN 420
	BEFORE AGING	AFTER AGING	
Elongation at Break (%)	550	500	ASTM D412
Force at break (N)	15	12	EN 455-2

ORDERING INFORMATION

Size	XS (5.5 – 6)	S (6.5 – 7)	M (7.5 – 8)	L (8.5 – 9)	XL (9.5 – 10)	XXL (10.5 – 11)	XXXL (11.5 – 12)
Product Code	93853060	93853070	93853080	93853090	93853100	93853110	93853120

Performance Standards and Regulatory Compliance



For additional information visit us at www.ansell.com, or call us at

Europe, Middle East & Africa Region

Ansell Healthcare Europe NV

T: +32 (0) 2 528 74 00
F: +32 (0) 2 528 74 01

Asia Pacific Region

MY: +603 8310 6688
apac.medical@ansell.com
CN: +86 21 38275000
infochina@ansell.com
JP: +813 5549 8151
info.medical.jp@ansell.com

North America Region

Ansell Healthcare Products LLC

T: +1 800 800 0444
F: +1 800 800 0445

Latin America & Caribbean Region

BR: +55 (11) 3356-3100
latam.medical@ansell.com
MX: +52 442 296 2050
latam.medical@ansell.com
CO: +57 1 288 3247
latam.medical@ansell.com

Australia

Ansell Limited

T: +61 1800 337 041
F: +61 1800 803 578

Russia

Ansell PYC
Ten. +7 495 258 13 16

Ansell, ® and ™ are trademarks owned by Ansell Limited or one of its affiliates. US Patented and US and non-US Patents Pending: www.ansell.com/patentmarking © 2025 Ansell Limited. All Rights Reserved.

Neither this document nor any other statement made herein by or on behalf of Ansell should be construed as a warranty of merchantability or that any Ansell product is fit for a particular purpose. Ansell assumes no responsibility for the suitability or adequacy of an end user's selection of gloves for a specific application.

